

April 19, 2012

Alaris Pump Missing Battery Alert

|
Please see the photo below. If this message presents when you power up the pump, it should be powered off and sent to Biomed for repairs.

The message does not have a red or yellow border as other fail or caution messages. Attempts to reset the pump can result in further messages or interruption of therapy.

